TECHNICAL NOTE

BIOIMPEDANCE MEASUREMENT OF SEGMENTAL FLUID VOLUMES AND HEMODYNAMICS

Leslie D. Montgomery, Ph.D.¹
Yi-Chang Wu, M.D.³
Yu-Tsuan E. Ku, M.S.¹
Wayne A. Gerth, Ph.D.²

Running Title: Fluid Shifts and Hemodynamics

Send correspondence and reprint requests to:

Leslie D. Montgomery
Lockheed Martin Engineering & Sciences Co.
Mail Stop 239-15
NASA Ames Research Center
Moffett Field, CA 94035

 Lockheed Martin Engineering & Sciences Co. NASA Ames Research Center Mail Stop 239-15 Moffett Field, CA 94035

- 2 Medical Research Assistant Professor of Physiology and Anesthesiology Department of Cell Biology, Division of Physiology Department of Anesthesiology Duke University Medical Center; Durham, NC 27710
- 3 Department of Cell Biology, Division of Physiology Duke University Medical Center; Durham, NC 27710

ABSTRACT

stresses.

Background: Bioimpedance has become a useful tool to measure changes in body fluid compartment volumes. An Electrical Impedance Spectroscopic (EIS) system is described that extends the capabilities of conventional fixed frequency impedance plethysmographic (IPG) methods to allow examination of the redistribution of fluids between the intracellular and extracellular compartments of body segments.

Methods: The combination of EIS and IPG techniques was evaluated in the human calf, thigh and torso segments of eight healthy men during 90 min of 6° head-down tilt (HDT).

Results: After 90 min HDT the calf and thigh segments significantly (P<0.05) lost conductive volume (8 and 4%, respectively) while the torso significantly (P<0.05) gained volume (approximately 3%). Hemodynamic responses calculated from pulsatile IPG data also showed a segmental pattern consistent with vascular fluid loss from the lower extremities and vascular engorgement in the torso. Lumped-parameter equivalent circuit analyses of EIS data for the calf and thigh indicated that the overall volume decreases in these segments arose from reduced extracellular volume that was not completely balanced by increased intracellular volume.

Conclusion: The combined use of IPG and EIS techniques enables noninvasive tracking of multi-segment volumetric and hemodynamic responses to environmental and physiological

Keywords: Bio-Impedance, Impedance Spectroscopy, Fluid redistribution, Hemodynamics, Head-down tilt, Intracellular volume, Extracellular volume

INTRODUCTION

Redistributions of body fluids between different body segments (i.e. legs, torso, and arms) and between the intra- and extracellular compartments within these segments, are important physiological features of shock and other clinical disorders [4] and of response and adaptation to various orthostatic and anti-orthostatic stresses, including microgravity [2,6-9,11,16,17]. These fluid redistributions affect cardiovascular function, water balance and perhaps skeletal muscle function [2,4,6,7] through physiological mechanisms that may be better understood with simultaneous characterization of both the fluid redistributions themselves and of associated changes in cardiovascular and hemodynamic parameters. Fixed frequency bioelectrical impedance plethysmographic (IPG) techniques have emerged as valuable noninvasive tools that provide information about overall segmental volumes and hemodynamic status [12,14,15,22]. However, these techniques cannot provide information about relative redistributions of fluids between the intra- and extracellular compartments of body segments. Electrical impedance spectroscopy (EIS), coupled with computer-aided equivalent circuit analysis, can measure such compartmental changes while retaining the other advantages of the older IPG methods.

Tissues are ionic conductors of electric current which, by virtue of their structural heterogeneity, exhibit dielectric relaxation phenomena that give rise to frequency dependent variations of such conductive impedance [1]. In the 1 - 150 kHz range, the resultant dielectric dispersions arise principally from the capacitive reactance of cell membranes. At low frequencies in this range, high cell membrane reactances prohibit current flux through cells so that tissue impedance is governed by properties of the extracellular fluid. At high frequencies membrane

reactance is negligible, thereby allowing current to pass through both extra- and in racellular spaces. Tissue impedance is then governed by the combined properties of the two compartments. Time series of tissue impedances measured as a function of frequency consequently embody tissue structural information that can illuminate changes in the relative distributions of fluid between the tissue intra- and extracellular compartments [1,10].

An EIS system that measures the *in-vivo* dielectric properties of tissue was used with conventional fixed frequency bioimpedance plethysmography to assess compartmental fluid redistribution and hemodynamic changes that occurred in human volunteers during short term 6° head-down tilt.

METHODS

System Overview

The combined impedance system used tandem operation of separate EIS and IPG instruments.

Electrical Impedance Spectroscopy The EIS system consisted of a Schlumberger Technologies, Inc. (New york, NY), Solartron 1260 Impedance/Gain-Phase Analyzer controlled via an IEEE-488 interface by a Digital Equipment Corporation (Houston, TX), VAXstation 3200 computer. The entire system (except a printer) was mounted in one full EIA standard cabinet rack on castors.

An impedance spectrum was obtained by measuring the voltage across a segment to sinusoidal electric current excitation at each of a series of discreet frequencies from 3 to 150 kHz. A signal generator and current amplifier provided the excitation signal which was passed

through the segment and terminated at a current input channel, where the amplitude and phase of the current were measured. Electric response of two segments in the excitation current path was measured across independent voltage input channels. Tetrapolar electrode configurations were used to minimize electrode impedance effects. System software was used to configure the analyzer for spectrum acquisition by setting appropriate values for all analyzer functions, including the excitation current and the frequencies to be swept. The analyzer was operated in differential mode on each of its two channels with the shields for all output and input leads floated from ground at the analyzer chassis. Shields of all leads were also brought to equal potential at a single point near the electrode ends of the leads.

The excitation current was fixed at about 0.5 mA or increased as a function of frequency from about 0.3 mA to a maximum of 5.0 mA (3.0 V max). The latter procedure maximizes signal-to-noise ratios and measurement accuracy where maximum safe drive currents decrease with decreasing frequency. As the measurement frequency increased during each sweep from about 2.5 kHz to 150 kHz, the drive current was adjusted upward using the log-log relationship: lnI = mlnf + b; where: I is the current at frequency f, and values of m and b were set conservatively according to data for the frequency dependence of the threshold for current sensation in the human thorax [5]. A minimum current of 0.3 - 1.0 mA was used when this relationship gave smaller values and a limit of 5.0 mA was imposed at higher frequencies. Data were passed in binary form from the analyzer to the computer for immediate processing and storage on disk or magnetic tape.

Impedance Plethysmography A tetrapolar, multi-channel impedance plethysmograph (UFI Inc., No. 2994, Morro Bay, CA) was used to measure baseline resistances (Ro) and pulsatile resistance changes (ΔR) in each monitored body segment. The IPG operated at a constant current, fixed excitation frequency of 50 kHz (0.1 mA, rms). A UFI Cardiotach (Morro Bay, CA) was used to monitor a lead I electrocardiogram (ECG). The seven analog outputs (6 impedance, 1 ECG) were each sampled at 200 Hz and recorded in digital form using a IBM-compatible personal computer with an 8-channel differential A/D convertor (Data Translation, Inc., DT-2811L, Marlboro, MA) running under control of a commercially available high-speed data acquisition system (DataQ, Inc., CODAS, Akron, OH).

System Data Acquisition

Each EIS impedance spectrum consisted of a series of discreet impedances (Z^*) computed from the measured voltage V^* and current I^* at each of the separate frequencies in the sweep, where:

$$Z^* = V^*/I^* = R + Xj$$
; (1)

where $j = \sqrt{-1}$ and R is the equivalent series resistance and X is the equivalent series reactance. A simple lumped-parameter equivalent circuit model representing idealized conductance paths through the body segment was fitted to each measured spectrum using a nonlinear least squares routine based on Marquardt's algorithm [13]. The circuit (Figure 1) models the segment as a uniform isotropic bidomain conductor [18] with an extracellular compartment having average resistance Re, and an intracellular compartment having average resistance Ri and average capacitance Cm. The value of Ri is governed by both membrane and cytoplasmic properties, while Cm is governed principally by membrane properties [1]. The complex admittance Y* of the circuit is given by

$$Y^* = \frac{1}{1} = \frac{1}{1} = \frac{1}{Ri}$$

$$Y^* = \frac{1}{1} = \frac{1}{1} = \frac{1}{Ri}$$

$$Z^* \quad Re \quad Ri \quad 1 + j\omega \tau^{o(1-\alpha)}$$
(2)

where: ω is the angular frequency given by $2\pi f$, τ^o is the time constant given by the product Ri·Cm, and $\alpha\pi/2$ is the angle between the real axis and a radius of the admittance locus passing through either of its two real axis intercepts. The $(1-\alpha)$ exponent in Eq. (2) is included to account for the typical failure of tissue impedance loci to be centered on the real axis. This behavior is consistent with the presence in tissue of a practical infinitude of parallel R-C elements each with different time constants with values distributed about a mean at τ^o [1,3,10]. By definition, α =0 when the center of the locus lies on the real axis. Increasing values of α from 0 towards unity indicate a widening of the distribution of time constants with increasing standard deviation of the distribution about τ^o .

Marquardt's algorithm was implemented using the norm of each observed impedance $|Z^*|$, and that of the corresponding fitted impedance $|Z^*|$. The algorithm adjusted the model parameters to minimize the sum of squares, (SS),

$$SS = \sum_{i=1}^{n} \{ |Z_{i}^{*}| - |Z_{i}^{*}| \}^{2},$$
(3)

for the impedances at the n different frequencies in each spectrum. The analytic components of the software were bundled to process run-time data passed from the impedance spectrum acquisition routine, or to read and process data files from earlier experiments thus providing identical output in either case. In the former mode the analyses were performed immediately after acquisition of each spectrum. Graphic display of the results afforded a means to track changes in measured and computed dielectric properties of each body segment throughout each run.

Impedance plethysmographic data were analyzed on a pulse-by-pulse basis using a custom interactive software package that included a graphic user interface to facilitate selection and specification of the impedance waveform landmarks shown in Figure 2. Times and differential resistances at these landmarks were used with basal resistances to calculate the following indices of segmental volume, blood flow and vascular compliance:

- (1) segmental conductive volumes (Ve) were calculated from segmental base resistances [15];
- (2) blood flow index, (BFI) is a function of the maximum amplitude of the impedance pulse, heart rate, and the basal resistance [14];

- (3) dicrotic index, (DCI) = B/A and is defined as the ratio of the amplitude of the pulse waveform at the height of the incisure (B) to the maximum pulse amplitude (A). DCI increases proportionally with general arteriolar compliance [22];
- (4) anacrotic index, (AI) = a/T and is defined as the ratio of the duration of the anacrotic phase of the pulse wave (a) to the duration of the entire cardiac cycle (T). This ratio is the relative systolic filling time of a given body segment during the cardiac cycle and reflects the compliance of the larger arteriolar vessels. AI decreases as the local arteriolar compliance increases [21]; and
- (5) pulse transit time (PTT) is the time interval in seconds between the onset of the ECG QRS complex and the onset of the impedance pulse waveform; i.e. the time required for the pressure profile of the cardiac pulse to be transmitted from the heart to the monitored segment.

 As the pulse conduction path becomes more rigid the pressure pulse is transmitted more quickly. Thus, PTT is an index of the overall vascular compliance of the body [20].

Subject Testing Procedure

Eight men (27 to 50 yrs) in good health provided informed consent and volunteered for the 90 min 6° head-down tilt (HDT) protocol. The investigation took place in the H. G. Hall Hypo/Hyperbaric Laboratory of Duke University School of Medicine and was fully approved by the Human Use Committees of both SRI International, Menlo Park, CA and Duke University School of Medicine. Each test sequence consisted of three successive periods when the subject was in the seated upright for 30 min, supine at 6° head-down tilt for 90 min, and again then

seated upright for 60 min. Except for standing to move from one position to the next, the subject was at rest and asked to minimize limb movement and muscle contraction throughout each run conducted at 21.0 ± 0.5 °C average ambient temperature where air movement, bright light, noise, and other stimulation were minimised.

Each subject was fitted with ECG electrodes (3M, Ag/AgCl Red Dot, St. Paul, MN) for bioimpedance monitoring of the left calf; thigh and torso. Calf pickup electrodes were placed laterally above the lateral malleolus of the fibula and below the fibular head. Thigh pickup electrodes were placed just above the lateral epicondyle of the femur and on the lateral aspect of the greater trochanter of the femur. Torso pickup electrodes were placed on the left iliac crest and the left clavicular line. Distances between the pickup electrodes for each segment were measured for calculation of segmental conductive volumes [15]. Drive electrodes were placed on the dorsum of the metatarsus and on the lateral aspect of the anterior superior iliac spine. An electrode on the back of the left hand was used in place of the latter for IPG monitoring to include the torso in the current path. Each subject was also instrumented with standard sternal and biaxillary ECG electrodes for monitoring heart rate.

EIS was used at five min intervals to characterize the dielectric properties of each body segment. Data acquisition for each impedance spectrum required about 70 s and measured 50 discreet frequencies distributed logarithmically from 3 to 150 kHz. At run elapsed times of 0, 40, 65, 115, 130 and 185 min. The electrodes were disconnected from the EIS system and connected to the IPG system. Two-minute periods of IPG measurements of baseline resistance

(Ro) and pulsatile resistance changes (ΔR) were then made of the calf, thigh, and torso body segments.

Statistical Analysis

Data were analyzed with the paired-t test of measured and calculated parameters.

RESULTS

Calf, thigh and torso conductive volume changes during the HDT period are shown in percent relative to the first value obtained after start of the HDT period; i.e., the value obtained at 40 min total elapsed time or 8-10 min after initial assumption of the HDT position (Figure 3). By minute 65 (approximately 33 min HDT) the calf segment had lost approximately 3% of its conductive volume (P<0.05), the thigh had lost about 2% volume (N.S.), and the torso had gained about 1.5% volume (N.S.). After 90 min HDT the volumes of all three body segments had changed significantly (P<0.05) relative to the first measurement: the calf and thigh had lost approximately 8 and 4% volume, respectively, and the torso volume had increased by approximately 3%.

Figures 4 and 5 present the grouped mean values (± S.E.) of the calculated segmental hemodynamic indices at the indicated elapsed times during the test protocol. Results are shown normalized and reduced to percent parametric change for each segment relative to the initial pretilt value (@ t=0). The grouped mean heart rate (Figure 4A) decreased approximately 10% (P<0.05) during the HDT period, and returned to pre-tilt values within minutes of assuming the

seated recovery position. The blood flow indices of the calf, thigh, and torso segmnts all increased (P<0.05) (Figure 4B) during the first 30 minutes of HDT, 50, 115 and 30% respectively. After 90 min HDT, the index for the thigh segment remained elevated, while indices for the calf and torso showed respectively greater tendencies to decrease toward their pretilt values.

The dicrotic index (Figure 5A) decreased (P<0.05) in the calf and thigh throughout the HDT period. In contrast, this index for the torso increased (P<0.05) during the first 30 min of HDT and decreased (P<0.05) to its pre-tilt value after 90 min HDT. The anacrotic index (Figure 5B) increased (P<0.05) between 20 and 40% in the calf and thigh but decreased (P<0.05) approximately 10% in the torso during HDT. Finally, segmental pulse transit times (Figure 5C) increased (P<0.05) in the lower body segments and decreased (P<0.05) in the torso during the HDT period. Changes in each of these separate segmental hemodynamic indices indicate that vascular compliance increased in the calf and thigh but decreased in the torso during HDT.

DISCUSSION

Results provide a detailed multi-segmental description of the volumetric and hemodynamic changes that occur during the initial 90 min of anti-orthostatic bed rest. This environmental stressor is widely used to simulate the cardiovascular effects of microgravity and has known effects on fluid distribution and cardiovascular performance [2,7,9,10,12]. In conformance with these known effects, cardiovascular, hemodynamic and volumetric parameters

provided by the present techniques exhibited systematic and dynamic changes during HDT that were reversible and showed patterns among the monitored segments that varied in accordance with the orthostatic dependence of the segments.

The IPG results indicate that the calf and thigh segments lost conductive volume while the torso gained volume during HDT. This pattern is typical of a general cephalad shift of fluids, where the volume gain in the torso represents a portion of the fluid displaced from the lower segments. Observed segmental volume changes are consistent with those determined in a similar fashion in other studies and shown to agree well with the volume changes determined from simultaneous anthropometric, capacitance band and strain-gauge plethysmographic measurements [15]. The agreement between the changes in mutual resistance calculated from the EIS results and the changes in segmental conductive volumes determined from the IPG data thus provides an important cross-validation of the two techniques: The segmental fluid shifts evident in the IPG results are also evident in the EIS results.

The EIS results provide additional insight into how the segmental intra- and extracellular compartments participate in the fluid shifts. Because the equivalent circuit model as presently applied ignores the inequalities of the transverse and longitudinal resistivities of muscle masses in the monitored segments, calculated compartmental resistance and capacitance values have only limited absolute significance [18]. However, for a given electrode orientation, fractional changes in these values accurately reflect actual changes in compartmental properties, provided that the processes leading to such changes uniformly affect all the monitored tissue. Thus, with assumption and maintenance of the HDT position, large initial increases in both calf and thigh Re reflect the exsanguination of capacitance vessels, while the more gradual increases in these

parameters continuing throughout HDT indicate interstitial fluid loss [10]. The latter was more pronounced in the calf than in the thigh, indicating, in accord with the different antiorthostatic stresses in the segments, that interstitial fluid loss in the thigh was smaller when compared to that in the calf. These results are consistent with previously reported decreases of interstitial and transcapillary pressures in the calf soleus muscle that continued throughout 8 hr of 5° HDT [7]. Upon resumption of the seated position, the fluid movement pattern reversed, with rapid return of the vascular volume in both segments followed by a gradual recovery of the respective interstitial volumes.

Changes in the Ri and Cm parameters illuminate changes in the volumes and ionic concentrations of the segmental intracellular spaces. The decreases in calf Ri that accompanied assumption and maintenance of the HDT position indicate that the average volume and/or electrolyte concentration of the intracellular space increased. The former interpretation is supported by the observed coincident increases of calf Cm. Increases in this parameter are plausibly interpreted to arise from decreases in the segmental average cell membrane thickness and increases in membrane area, such as would accompany increases in average cell volume. The indication that cells in the calf swell when the body is under anti-orthostatic stress is in accord with evidence that calf interstitial osmotic pressure decreases during HDT [7]. Similar but attenuated indications were obtained for the compartmental responses in the thigh.

The EIS results consequently augment the IPG results by providing evidence that HDT-induced losses of conductive volume from the calf and thigh arise from rather large losses of extracellular fluid that are not completely balanced by cellular imbibition of fluid. Similar conclusions have been forwarded by other workers for humans during chair rest and water

immersion, where significant decreases of serum osmobility were observed [6], and during horizontal bed rest [8]. Hargens, et al., [7] also reported increases of cellular area in connective tissue of soleus muscle biopsied from human subjects after 8 hr HDT, although a net intracellular to extracellular fluid shift -- opposite to that presently indicated -- was concluded to have occurred on the basis of increased plasma or urine potassium and decreased creatinine urinary excretion. Decreases of serum osmolality, the most probable cause of intracellular fluid accumulation, evidently occur in astronauts during spaceflight [11]. Notwithstanding the controversy about the direction of compartmental fluid shifts induced by anti-orthostasis, the presently indicated shifts are surprisingly large. The relative diminution of the shifts in the thigh with respect to those in the calf suggest that the net whole-body shift must be smaller, but the mechanism for such segmental differentiation is unknown. Further validation of the method is required to determine to what extent the effects might be quantitatively overestimated as a result of nonuniformities of current flow in the monitored segments and changes therein that could be caused by changes in the relative fluid contents of the composite tissues.

As observed for the segmental volumetric responses, the hemodynamic parameters calculated from the pulsatile IPG data also showed segment-specific responses to changes in orthostatic stress. Both the calf and thigh manifested hemodynamic responses to HDT that would have tended to minimize intravascular fluid loss from those segments. Decreases in the dicrotic index, and increases in the anacrotic index and pulse transit time indicate that the overall calf and thigh vascular compliance increased during HDT. These changes tended to be more pronounced in the calf than in the thigh, reflecting the same orthostatic dependence evident in the segmental fluid shifts. Conversely, hemodynamic parameters in the torso exhibited changes

during HDT that would have tended to minimize vascular engorgement in this segment. Thus, segmental hemodynamic changes during HDT appear to have been coordinated to limit or reduce the transfer of intravascular fluid from the legs to the torso and head. This counterregulatory pattern is consistent with that found by other investigators in human subjects under similar anti-orthostatic stresses [9,12,19], and probably arises from both active changes in arteriolar tone and passive volume-related changes in compliance.

Segmental blood flow changes during HDT were generally consistent with the indicated changes in vascular compliance, reflecting the overall efficacy of the latter at redistributing an increased central venous return. Blood flows increased and remained elevated in the calf and thigh segments during the HDT period. Blood flow in the torso increased at the beginning of HDT, but tended to decrease toward the pre-tilt value during the latter stages of HDT. Heart rate remained decreased throughout HDT. Given the increased torso blood flow, this result indicates that cardiac stroke volume increased during HDT. Results are generally in accord with earlier indications that cephalad shifts of blood during HDT cause increased end-diastolic volume and cardiac output, with the latter causing segmental blood flow increases [2,16,17]. The presently observed tendency for return of the cardiovascular parameters to pre-tilt values after an initial HDT-induced transient is also consistent with results of other studies, although the present time course for such adaptation appears to be longer than in other work [2].

CONCLUSIONS

The combination of fixed and swept frequency bioimpedance monitoring has provided a detailed characterization of the human body's regional volume, hemodynamic and blood flow

responses to a short term head-down tilt. This study provides new information regarding the extent and temporal nature of the fluid redistribution and segmental hemodynamics of the lower limbs and torso during HDT. Vascular dynamics in the legs were found to react in such a way as to inhibit the transfer of fluid volume out of the legs while those of the torso responded to minimize the transfer of fluid into the torso during HDT. The use of EIS provides additional information regarding the interaction of the intra- and extracellular volumes of the legs during HDT. Observed responses are within the ranges of those reported from other studies, while regional patterns of the responses indicate that studies of this type should include examination of the relevant parameters to at least the segmental level of anatomic resolution. The present techniques provide such resolution augmented by a high level of temporal resolution using only simple superficial electrodes. Occlusion cuffs, displacement fluids, confining garments or invasive procedures are not required, making the methods suitable for applications under either laboratory or field operational conditions. These techniques afford a useful noninvasive means to monitor a variety of physiological parameters concerning the physiology of fluid control in aerospace and clinical environments.

ACKNOWLEDGEMENT

This research was supported in part by a contract from the U.S. Navy Bureau of Medicine, Contract No. N00014-87-C-0166. The authors thank Dr. Richard W. Montgomery of Management Analytical Associates, Inc., Corvallis, OR, for assistance in software development and Dr. Marc Maddou of SRI International, Menlo Park, CA, for technical advice during the initial phases of this work.

REFERENCES

- 1. Ackmann JJ, Seitz MA. Methods of complex impedance measurement in biologic tissue. CRC Crit Rev Biomed Eng 1984; 11:281-311.
- 2. Butler GC, Xing H, Hughson RL. Cardiovascular response to 4 hours of 6° head-down tilt or of 30° head-up tilt bed rest. Aviat Space Environ Med 1990; 61:240-6.
- 3. Cole KS, Cole RH. Dispersion and absorption in dielectrics. I. Alternating current characteristics. J Chem Phys 1941; 9:341-52.
- 4. Gauer OH, Henry JP, Behn C. The regulation of extracellular fluid volume. Ann Rev Physiol 1970; 32:547-95.
- 5. Geddes LA, Baker LE. Hazards in the use of low frequencies for the measurement of physiological events by impedance. Med Biol Eng 1969; 7:289-96.
- 6. Greenleaf JE, Shavartz E, Kravik S, Keil LC. Fluid shifts and endocrine responses during chair rest and water immersion in man. J Appl Physiol 1980; 48:79 88.
- Hargens AR, Tipton CM, Gollnick PD, Mubarak SJ, Tucker BJ, Akeson WH. Fluid shifts and muscle function in humans during acute simulated weightlessness. J Appl Physiol 1983; 54:1003-9.
- 8. Johnson PC, Driscoll TB, Carpentier WR. Vascular and extravascular fluid changes during six days of bedrest. Aerospace Med 1971; 42:875-8.
- 9. Kabesheva TA, Kopanev SV, Panfervoa NY, Zavadovshiy AF. Vascular mechanisms of adaptation to antiorthstatic position. Kosmich Biol Aviak Med 1985; 19:35-9.

- 10. Kanai II, Katsuyuki S, Haeno M. Electrical measurement of fluid distribution in human legs:

 Estimation of extra- and intracellular fluid volume. J Microwave Power, 1983; 18:233-43.
- 11. Leach CS. Fluid control mechanisms in weightlessness. Aviat Space Environ Med 1987; 58(9,Suppl.):A74-9.
- 12. Maksimov DG, Domracheva MV. Changes in central and peripheral hemodynamics during prolonged anti-orthostatic hypokinesisa as weightlessness models. Kosmich Biol Aviak Med 1976; 5:52-7.
- Marquardt DW. An algorithm for least-squares estimation of nonlinear parameters. J Soc
 Indust Appl Math 1963; 11:431-41.
- 14. Montgomery LD, Hanish HM, Marker RA. An impedance device for study of multisegmental hemodynamic changes during orthostatic stress. Aviat Space Environ Med 1989; 60:116-22.
- Montgomery LD. Body volume changes during simulated weightlessness: An overview.
 Aviat Space Environ Med 1987; 58(9,Suppl.):A80-5.
- 16. Nicogossian AE, Parker Jr. JF. Space physiology and medicine. National Aeronautics and Space Administration 1982; NASA SP-447:141-52.
- 17. Nixon JV, Murray RG, Bryant C, Johnson Jr. RL, Mitchell JH, Holland OB, Gomez-Sanchez C, Vergne-Marini P, Blomquist CG. Early cardiovascular adaptation to simulated zero gravity. J Appl Physiol 1979; 46:541-8.
- 18. Plonsey R, Barr RC. A critique of impedance measurements in cardiac tissue. Annals Biomed Eng 1986; 14:307-22.

- 19. Skagen K. Sympathetic reflex control of blood flow in human subcutaneous tissue during orthostatic maneuvers. Danish Med Bull 1983; 30:229-41.
- 20. Strandness DE, Sumner DS. Hemodynamics for Surgeons. New York: Grune and Stratton 1975: 187-96.
- 21. Usochev VV, Shinkarevskaya IP. Functional changes in systemic and regional (intracranial) circulation accompanying low acceleration. Kosmich Biol Aviak Med 1973; 19:59-64.
- 22. Yarullin KK, Krupina TN, Vasil'yeva TD, Buyvolova NN. Changes in cerebral, pulmonary, and peripheral blood circulation. Kosmich Biol Aviak Med 1972; 6:33-9.

FIGURE LEGENDS

- 1. Equivalent circuit used to interpret measured segmental impedance spectra.
- 2. Impedance plethysmographic pulse waveform with landmarks for calculation of hemodynamic parameters.
- 3. Grouped mean segmental conductive volumes vs elapsed time. An asterisk (*) indicates that the value differs significantly (P<0.05) from the initial head-down tilt value. A plus (+) indicates that the value differs significantly (P<0.05) from that at 65 min elapsed time.
- 4A. Grouped mean heart rates vs elapsed time. An asterisk (*) indicates that the value differs significantly (P<0.05) from that during the pre-tilt seated position. A plus (+) indicates that the value differs significantly (P<0.05) from the preceding value.
- 4B. Grouped mean blood flow indices vs elapsed time. Symbols indicate significant differences as in Figure 4A.
- 5A. Grouped mean segmental dicrotic indices vs elapsed time. Symbols indicate significant differences as in Figure 4A.
- 5B. Grouped mean segmental anacrotic indices vs elapsed time. Symbols indicate significant differences as in Figure 4A.
- 5C. Grouped mean pulse transit times vs elapsed time. Symbols indicate significant differences as in Figure 4A.





Figure 1.



α = DURATION OF ANACROTIC PHASE OF EACH PULSATILE WAVE (s)

T = DURATION OF TOTAL CARDIAC CYCLE (s)

A = MAXIMUM LEVEL OF THE IMPEDANCE WAVE (mm)

B = AMPLITUDE AT THE POSITION OF THE DICROTIC NOTCH (mm)

C = AMPLITUDE AT THE LEVEL OF THE PEAK OF THE DICROTIC WAVE (mm)

